



Spectro UV-VIS Double PC 8 Auto Cell Scanning Spectrophotometer

Model UVD-3200

Software Specifications

Monoprocessor Built-in Application:

Photometric Measurement: Measuring transmittance or absorbance at the current wavelength together with K factor calculations.

Spectrum Scan: Carrying out scanning of transmittance or absorbance on the selected wavelength range together with peak-pick module.

Quantitative Determination: Regression of standard curves and direct determination concentration of samples.

PC Windows Application Software (RS-232 Interface) to link Spectro to computer and printer:

Photometric Measurement: Measuring the photometric values at 1-10 wavelengths together with mathematical calculations according to entered quotations.

Spectrum Scan: Producing Wavelength scans within the operating parameters on samples together with powerful data handling facilities.

Quantitative Determination: Determination of unknown concentration with methods of 1-3 wavelength quantitation, together with fitting of calibration curve of 1st ~ 4th order.

Kinetics: Recording curves of changing photometric values of samples against timecourse at the selected wavelengths together with powerful data handling facilities.

Output: With the Windows clipboard, the measured data and graphics can be copied to other applications software for reports.

Technical Specifications

Wavelength range:	190 nm – 1100 nm	Baseline Stability:	0.0008Abs/h (1/2 hr warmup, 1nm bandwidth, at 500 nm)
Spectral Bandwidth:	0.5, 1.0, 2.0 and 5.0 nm (UVD-3200 4 steps) 2.0nm (UVD-3000)	Slew Rate of Wavelength:	3600nm/min
Resolution:	0.1 nm	DNA/RNA Measurement:	Results Printout: Printing of measured data by using any Printer with Parallel Port connection available.
Straylight:	>2.10Abs (200nm)	Mainframe:	Compact and standalone spectrophotometer mainframe
Wavelength Accuracy:	±0.3 nm (with automatic wavelength correction)	Light Source:	Socket Deuterium Lamp and Socket Tungsten Halogen Lamp
Wavelength Reproducibility:	±0.2 nm	Detector:	Double Beam
Photometric System:	Double beam optical system	Sample Chamber:	Automatic eight-cell sample
Photometric Method:	Transmittance, absorbance, energy, concentration	Display	Liquid Crystal Display (LCD 320 - 240 dot matrix)
Photometric Range:	-0.3~3.0 Abs (0~200%T)	Keypad:	Touch soft keys.
Photometric Accuracy:	± 0.002Abs (0 ~ 0.5Abs), 0.004Abs (0.5 ~ 1.0Abs), ± 0.3% T (0 ~ 100% T)	PC Interface:	PC Interface: RS-232
Photometric Reproducibility:	0.001Abs (0~0.5 Abs), 0.002Abs (0.5~1.0Abs), T (0 ~ 100% T) 0.15%	Size:	22" x 16" x 10"
Photometric Display:	-9999 ---- 9999	Weight:	55 Lb
Photometric Noise:	± 0.001Abs (500 nm, 1 nm bandwidth, 0Abs)		
Scanning Speed:	1400nm/min		
Baseline Flatness:	±0.0015Abs (200 nm. ~1100 nm)		